



Alcohol Detector AL6000

Description

The AlcoMate Prestige is intended to measure alcohol in human breath. The accuracy of this device has been established at a Blood Alcohol Concentration of 0.008 and 0.032. Accuracy at a Blood Alcohol Concentration greater than 0.032 has not been evaluated. Accuracy testing was completed according to DOT regulations.

Method Of Measure: Breath Sample To Bac

Range: 0.00% To 0.40%

Accuracy: +/-0.01% Up To 0.10% Bac

Sensor: Oxide Semiconductor

Power: Two 1.5v Aa Alkaline Batteries (Included)

Dot Approved: Yes

Operating Temperature: 10 – 40 Degrees Celsius

Warm Up: Less Than 1 Minute

Display: 3-digit (0.00)

Size (Mm): 125 (Height) X 58 (Width) X 41 (Thickness)

Weight: 117 G

Sensor Modules: Sensor Modules Should Be Replaced Every 500 Tests Or At Least Once Yearly (Whichever Comes First)

Warranty:- 6 Months



Additional information

Weight :

0.23 kg

KTI Draft