



KT AUTOMATION PRIVATE LIMITED

CIN: U52100GJ2021PTC121843

Reg. Office: D-33 Kamalanjali Complex, above HDFC Bank, Akota,
Vadodara, Gujarat, India – 390020
(M):9227788388

Website: ktindia.net Email ID: contact@ktindia.net

Auto Inhale: No need to touch and fast detect

SOS: with strong LED and Vibration

Whistle: Powerful whistle

Red LED

Magnet Mount: Fixed onto vehicle and other metal

Rechargeable Battery: 3000mAh Lithium-ion battery



Non-Contact Breath Alcohol Analyzer model MS10QC

The parameter name	Specification
Brand	Mangal
Sensor Type	Semiconductor sensor
Measuring Range	0-400mg/100L BAC
Sensitivity/Differentiation Rate	5mg/100mL / 1mg/100mL
Indication Error	5cmThe distance concentration is 50mg/100 mL hours \pm 10mg/100mL
Sampling Method	Air blow induction / manual detection
Continuous Measurement Time	5 seconds to 15 seconds
Value Out Time	3 Seconds
Operating Temperature	-5~45 degree
Return to zero time	5 seconds to 15 seconds
Data Storage	30000 records
Operation Method	4 buttons, the warning light and lighting can be operated independently in the shutdown stat
Accessibility	LED High-bright white light illumination, high decibel whistle, red and blue waring lights, voice broadcast
Display Screen	1.3-inch 240*240 full-color TFT LCD screen
Supply Electricity	18650-3000mAh lithium battery
USB	TYPE-CUSB
Charging Voltage	5V/2A
Host size	360mm*45mm*43mm
Weight	260g